# NEMA SCA&I Cardiovascular Fluoroscopic Benchmark Phantom

#### Model 901



## **DESIGNED TO MEET NEMA XR 21**

The CIRS Model 901 NEMA-SCA&I phantom was designed to evaluate and standardize catheterization image quality. It's the result of collaborative efforts between the Society for Cardiac Angiography and Interventions and the National Electric Manufacturers Association.

Our Model 901 is manufactured from PMMA with X-ray absorption properties similar to soft tissue at diagnostic energies. It contains a variety of static and dynamic test targets for objective assessment of resolution, motion unsharpness and radiation exposure. The sectional design allows for configuration in a wide range of thicknesses from 5 to 30 cm simulating PA thicknesses from infants to large adult patients.


The phantom is ideal for routine assessment of the entire imaging system.

#### Features

- Designed to meet NEMA XR 21
- Simulates coronary arteries for infant through large adult
- Evaluate spatial resolution, motion unsharpness and exposure for the entire imaging system

900 Asbury Ave • Norfolk, Virginia 23513 • USA Tel: 800.617.1177 • 757.855.2765 • Fax: 757.857.0523 *WWW.CIRSINC.COM* 





Model 901-01, 901-03, and 901-13. Model 901-13 Rotating Spoke Target is included with phantom, but can also be purchased separately.

## **SPECIFICATIONS**

DIMENSIONS:	40.6 cm x 25.4 cm x 25.4 cm
	(16" x 10" x 10")
PHANTOM WEIGHT:	20.4 kg (45 lb)
MATERIALS:	PMMA

#### **MODEL 901 INCLUDES**

Specifications subject to change without notice.

Publication: 901 DS 031620


QTY	PART NUMBER	DESCRIPTION
1	901-01	Central Target Assembly
1	901-02	Working Thickness Range (WTR) Plate A
1	901-03	WTR Plate B
1	901-04	WTR Plate C
3	901-05	WTR Plate D
1	901-06	WTR Plate E
4	901-07	Blank Plate with alignment parts
1 (ea.)	901-08/ 08A	Field Size Plate
1	901-09	Alignment Target for Test Stand
1	901-10	Alignment Cross for Test Stand
1	901-11	Alignment Target for small base
1	901-12	Alignment Cross for small base
1	901-13	Rotating Spoke Target (115 VAC 60 Hz)
1	901-14	Test Stand
1	901-15	Small Base
1	901-16	3 mm thick lead plate with laminate
1	901-17	2 mm thick copper plate with laminate
1	901-CS	Carry Case
35	-	Alignment Pins


### **OPTIONAL ACCESSORIES**

PART NUMBER	COMPONENT DESCRIPTION	
901-C	Step Down Transformer (220V)	
901-19	Optional Artery Slot Block for Model 901	
903-06*	Artery Block with lodine (15 mg/ mL)	
903-07*	Artery Block with lodine (150 mg/ mL)	

\*Must purchase Model 901-19 Optional Slot Block to use this accessory with Model 901 NEMA SCA&I Phantom.

#### ARTERY BLOCK WITH IODINE (903-06 & 903-07)





UĮ

CITERED.